

## TMC 011-ANAEROBIC BOX -L

### INTENDED USE

This Anaerobic box is used for cultivation of anaerobic microorganisms without the addition of water or use of a catalyst

### DESCRIPTION

For Laboratories routinely testing for microbes requiring a CO<sub>2</sub> rich, reduced O<sub>2</sub> atmosphere, the Anaerobic box-L provides a convenient solution. The Anaerobic box-L is lightweight yet robust. The uncomplicated design makes it easy to maintain and load with a compartment for both the gas kit and also indicator, minimizing the possibility of malfunction. The shape of the Anaerobic box-L also makes it easy to store and will save space in the incubator over more conventional round jars with bulky lids that include valves and gauges.

### DIMENSIONS AND CAPACITY













S. No	Description	Size	Capacity	Material
1	Transparent unbreakable poly box, with easy-close lid	28.5cm x 22cm x 11cm	7.0 Litres	Polycarbonate

### ACCESSORIES REQUIRED BUT NOT PROVIDED.

Anaerobe Gas Pack 3.5 Ltr. (TMC 001D) 2 Packs and Anaerobe Indicator Tablet (TMC 001C) 1 Pack

### CLEANING AND DECONTAMINATION

The disc dispenser should be cleaned using distilled water, if necessary with a few drops of liquid soap added to it. Finally, rinse with clean distilled water and dry.

 <b>GMP</b> Good Manufacturing Practices Certified	 <b>IVD</b> For In Vitro Diagnostic Use	 <b>QTY.</b> Quantity	 <b>LOT/ B. NO.</b> Lot / Batch Number	 <b>REF</b> Catalogue Number	 <b>Manufacturer</b>
 <b>Temperature Unit</b>	 <b>EC REP</b> MedNet GmbH Barkhausen 10, 48163 Münster, Germany Authorized Representative	 <b>CE</b> European Conformity	 <b>QR Code</b>	 <b>Consults Instructions for Use</b>	 <b>Best Before</b>

**NOTE:** Please consult the Material Safety Data Sheet for information regarding hazards and safe handling Practices.

**\*For Lab Use Only**

Revision: 10<sup>th</sup> June. 2020.

