

## TS 009 – CAMPYLOBACTER SUPPLEMENT-III (Skirrow)

### INTENDED USE

For selective isolation of *Campylobacter* species.

### COMPOSITION

Ingredients	Concentration
Vancomycin	5mg
Polymyxin B sulphate	1250IU
Trimethoprim	2.500mg

(Per vial sufficient for 500 ml medium)

### Applied in Medium

TM 041 - BLOOD AGAR BASE No. 2  
TM 514 - BRUCELLA AGAR BASE  
TMV 514 - BRUCELLA AGAR BASE (VEG.)  
TM 680 - BRUCELLA AGAR BASE, MODIFIED  
TM 056 - CAMPYLOBACTER AGAR BASE  
TM 071 - COLUMBIA BLOOD AGAR BASE  
TMV 071 - COLUMBIA BLOOD AGAR BASE (VEG.)  
TM 2044 - COLUMBIA BLOOD AGAR BASE W/ 1% AGAR  
TM 1842 - BLOOD AGAR BASE NO. 2 (ISO 11290-1:2017)  
TM 2011 - BLOOD AGAR BASE NO. 2 WITH 1.2% AGAR













### INSTRUCTION FOR USE

Rehydrate the contents of one vial aseptically with 2 ml sterile distilled water and mix well to dissolve. Avoid frothing of the solution. Aseptically add the rehydrated contents to 500 ml of sterile, molten, cooled (45-50°C) TM 071 - COLUMBIA BLOOD AGAR BASE / TMV 071 - COLUMBIA BLOOD AGAR BASE (VEG.) / TM 2044 - COLUMBIA BLOOD AGAR BASE W/ 1% AGAR / TM 041 - BLOOD AGAR BASE No. 2 / TM 1842 - BLOOD AGAR BASE NO. 2 (ISO 11290-1:2017) / TM 2011 - BLOOD AGAR BASE NO. 2 WITH 1.2% AGAR / TM 056 - CAMPYLOBACTER AGAR BASE with 5-7% v/v lysed defibrinated horse blood TM 514 - BRUCELLA AGAR BASE / TMV 514 - BRUCELLA AGAR BASE (VEG.) / TM 680 - BRUCELLA AGAR BASE, MODIFIED with 5% inactivated horse serum. Mix gently and pour into sterile petri plates.

### STORAGE

Vials should be stored in sealed container at 2-8°C.



 GMP Good Manufacturing Practices Certified	 IVD For In Vitro Diagnostic Use	 QTY. Quantity	 LOT/ B. NO. Lot / Batch Number	 REF Catalogue Number	 Manufacturer
 Temperature Unit	 EC REP Authorized Representative <small>MedNet GmbH Südstraße 10 48143 Aachen, Germany</small>	 European Conformity	 QR Code	 Consults instructions for Use	 Best Before

**NOTE:** Please consult the Material Safety Data Sheet for information regarding hazards and safe handling Practices.

**\*For Lab Use Only**  
Revision: 9 March., 2022